

地址 (Address): 湖北省武汉市东湖新技术开发区茅店山中路二号
网址 (Web site): <http://www.hbjl.gov.cn>

邮编 (Post Code): 430223
电话 (Tel): 027-81928136
传真 (Fax): 027-81925137

CUC

REGISTRATION NO. 002

校准证书

CALIBRATION CERTIFICATE

证书编号:

Certificate No.

委托方名称

Customer

厦门泓宇智能科技有限公司

委托方地址

Address

厦门市湖里区泗水道599号406室B单元

相互承认协议

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards.

- 本院是政府计量行政部门依法设立法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

- 本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

- 本次校准的技术依据（名称、代号）

Reference documents for the Calibration (Name 、 Code)

参照: JJG 496-2016 工频高压分压器

三

- 本次校准所使用的主要计量标准

校准数据/结果

Data/Result of Calibration

一、外观检查: 正常

1、高电压测量: AC

被测显示值 (kV)	实测标准值 (kV)	测量不确定度 U (kV), $k=2$
11.3	10.71	0.10
20.6	20.11	0.20
30.9	30.89	0.30
40.9	40.77	0.40
50.3	50.10	0.50
60.1	59.89	0.60

以下空白