



湖北省计量测试技术研究院

中国校准服务联盟成员



- This body is an institute of legal verification.

Authorization certificate number of the national metrological verification institution:

Laboratory accreditation certificate number of China National Accreditation Service for Conformity Assessment(CNAS): No.

校准数据/结果

Data/Results of Calibration

电压测量: AC	
被检显示值 (kV)	实测标准值 (V)
10	20.66
20	40.94
30	59.86
40	80.46
50	101.26
测量不确定度: $U=0.05$	

以下空白

卡研光院