

校准证书

CALIBRATION CERTIFICATE

证书编号:
Certificate No.委托方名称
Customer

安徽省长江计量所

委托方地址
Address

安徽省合肥市

器具名称
Name Of Sample

电流互感器现场测试仪

制造厂商

武汉特高压电力科技有限公司

Manufacturer

型号规格

HGQL-H

Model/Type

器具编号

141588

No Of Sample

校准专用章

Stamp



校准日期

Calibration date

2022年 08月 09日
Y M D

建议再校准日期

Recommended recal date

2023年 08月 08日
Y M D证书批准人
Approved by核验员
Checked by

校准员

Calibrated by

*本校校准所用的测量装置均溯源至保存量值, 国家计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基准及国家计量研究院出具的校准和测量证书相互承认协议”。

The calibration is traceable to national primary standards maintained in

湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

证书编号: 号

C

● 本院(中心) 严格按照法定计量检定规程, 量值溯源和校准, 是国家计量委员会(CNAS) 认可实验室

This body is an institute of legal verification, it is also a accreditation laboratory accredited by China National Accreditation Service for Conformity Assessment(CNAS).

● 国家法定计量检定机构授权证书号: (国) 法计(2012) 01023, (国) 法计(2012) 01042

Authorization certificate number of the national legal metrological verification institution:

中国合格评定国家认可委员会实验室: No. L0544

Laboratory accreditation certificate number of China National Accreditation Service for Conformity Assessment(CNAS): No. L0544

● 本次校准的技术依据(名称、代号)

Reference documents for calibration:

11.2

11.2.1

11.2.1

1000A/5A	199.9	0.1
----------	-------	-----

五、PT 变比测量:

标准示值	实测值	测量不确定度: $U(k=2)$
2000V/100V	20.00	0.01
3000V/100V	30.00	0.01
5000V/100V	50.01	0.01
6000V/100V	59.96	0.02
10000V/100V	99.99	0.02

六、导纳测试:

标准示值		实测值		测量不确定度 $U(k=2)$	
电导值(mS)	电纳值(mS)	电导值(mS)	电纳值(mS)	电导值(mS)	电纳值(mS)
0.100	0.000	0.098	0.001	0.001	0.001
1.000	0.000	0.989	0.003	0.005	0.005
10.00	0.000	9.88	0.02	0.01	0.01

以下空白