



- 本院是政府计量行政管理部门依法设立的法定计量检定机构
This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

- 本院质量管理体系符合ISO/IEC17025标准的要求。
The quality management system for laboratory complies with ISO/IEC 17025 standards.

- 本次校准的技术依据（名称、代号）
Reference documents for the Calibration (name & code)
参照: JJG1005—2005 电子式绝缘电阻表检定规程 Verification Regulation of Electronic Insulating Resistance Meters

- 本次校准使用的计量标准器具
Main standards of measurement used in the Calibration

设备名称

型号/编号

证书号/有效期

地址: 本城光谷基地B211室

原始记录编号:

校准数据/结果

Data/Results of Calibration

选择电压 (V)		选择电压 (V)	
0.1k		0.25k	
标准实际值	被校显示值	标准实际值	被校显示值
$U_{rel} (k=2)$		$U_{rel} (k=2)$	
1.000M Ω	0.97M Ω	1.000M Ω	1.04M Ω
2.000M Ω	1.97M Ω	2.000M Ω	2.04M Ω
5.000M Ω	4.95M Ω	5.000M Ω	5.04M Ω
10.00M Ω	9.90M Ω	10.00M Ω	10.04M Ω

500.0M Ω	0.495G Ω		500.0M Ω	0.500G Ω	1.2%
1000M Ω	1.025G Ω		1000M Ω	0.986G Ω	
2000M Ω	2.02G Ω	2.2%	2000M Ω	1.93G Ω	2.2%
选择电压 (V)		0.5k	选择电压 (V)		1k

